



GOVERNMENT OF KARNATAKA
KARNATAKA STATE MEDICAL SUPPLIES CORPORATION LIMITED

Sheshadri road, Opp S.J.P College, K.R Circle, Bengaluru.

Phone: +91-80-23486221 Fax: +91-80-23281477

Email: helpdesk.kdlws@gmail.com, ad.kdlws@gmail.com

No. HFW/KSMSCL/COVID-19/Oxygen Concentrator/34/2021-22

Dated: 22/05/2021

CORRIGENDUM

INVITATION FOR QUOTATIONS (IFQ) FOR SUPPLY AND INSTALLATION MEDICAL EQUIPMENTS FOR KCG FOR COVID-19 PANDEMIC.

1. TERMS AND CONDITION:

Sl. No	Existing Clause	Changed to
1	Proof of 30% Performance shall be submitted in the form of Supply Order/ Purchase Order & Invoice Copies.	Deleted. Read as copy of the purchase order and Invoices.
2	End user certificate to be submitted by bidder.	Deleted.

2. TECHNICAL SPECIFICATION:

Sl. No	Existing Clause	Changed to
1	The Concentrator should be able to produce minimum 10 LPM or flow rate of the concentrator should be 0.5-10 LPM.	Flow rate of the concentrator should be 0.5-10 LPM.
2	Should have low operating noise (<55dB).	Should have low operating noise (≤55dB).
3	Purity of output should be 93±3 at 10LPM oxygen purity should be equivalent to 93% ±3%.	At 10 LPM oxygen purity should be equivalent to 93% ±3%.

Sd/-

**Managing Director,
KSMSCL, Bengaluru**