



GOVERNMENT OF KARNATAKA  
KARNATAKA STATE MEDICAL SUPPLIES CORPORATION LIMITED

No.1, Dr. SiddaiahPuranik Road, KHB Colony, Magadi Road, Bangalore – 560079  
Phone: +91-80-23486221 Fax: +91-80-23281477

Email: [helpdesk.kdlws@gmail.com](mailto:helpdesk.kdlws@gmail.com), [ad.kdlws@gmail.com](mailto:ad.kdlws@gmail.com)

No.HFW/KSMSCL/GH-IFQ/06/2021-22

Date: 28-4-2021

INVITATION FOR QUOTATIONS (IFQ) FOR SUPPLY OF .....

**Corrigendum**

**Sub:** Invitation of quotation for supply of Drugs, Chemicals and Miscellaneous.

- Ref:**
1. Quotation Notification No. HFW/KSMSCL/COVID/ N 95 Mask/09/2021-22 Dated: 26.04.2021.
  2. Quotation Notification No. HFW/KSMSCL/ COVID/ TAB/13/2021-2022, Dated:27/04/2021.
  3. Quotation Notification No. HFW/KSMSCL/COVID/PPE KIT/10/2021-22, Dated: 26.04.2021
  4. Quotation Notification No. HFW/KSMSCL/COVID/T- Layer/08/2021-22, Dated: 26.04.2021
  5. Quotation Notification No. HFW/KSMSCL/GH- IFQ/06/2021-22, Dated: 26.04.2021.
  6. Cabinet Approval Date: 26.04.2021

\*\_\*\_\*\_\*

With Respect to Procurement of Essential Drugs/Chemicals/Other Items to deal with Public Health Emergency of International Concern (Covid-19), following amendments are made to the following Quotation Published.

1. Quotation Notification No.HFW/KSMSCL/COVID/N-95 Mask/09/2021-22 Dated: 26.04.202,

I.

Drug Name and Specification published in quotation	Amended as
<b>N-95 Mask</b> <ul style="list-style-type: none"><li>• Shape that will not collapse easily.</li><li>• High filtration efficiency.</li><li>• Good breathability, Quality Compliant with standards for medical N95 respirator:<ol style="list-style-type: none"><li>a. NIOSH N95, EN 149FFP2 or BIS.</li></ol></li></ul> Fluid resistance: Minimum 80 mmHg Pressure based an ASTM F1862, ISO 22609 or equivalent. Quality compliant with standards for particular respirator	<b>N-95 Mask</b> <ul style="list-style-type: none"><li>• Shape that will not collapse easily.</li><li>• High filtration efficiency.</li><li>• Good breathability, Quality Compliant with standards for medical N95 respirator:<ol style="list-style-type: none"><li>a. NIOSH N95, EN 149FFP2 or BIS.<ul style="list-style-type: none"><li>• Fluid resistance: Minimum 80 mmHg Pressure based an ASTM F1862, ISO 22609 or equivalent.</li><li>• Should have Adjustable Head Band.</li></ul></li></ol></li></ul>

that can be worn with full-face shield

- Quality compliant with standards that can be worn with full-face shield

II. Quantity of N-95 Mask Amended as (2,00,000+6,00,000) **8,00,000.**

2. Quotation Notification No. HFW/KSMSCL/COVID/ TAB/13/2021-2022,  
Date:27/04/2021.

i.

Drug Name and Specification published in quotation	Amended as
Tab. Zinc Sulphate Dispersible 50mg	Tab. Zinc Sulphate 50mg

3. Quotation Notification No. HFW/KSMSCL/COVID/PPE KIT/10/2021-22,  
Date: 26/04/2021.

i.

Drug Name and Specification published in quotation	Amended as
<ul style="list-style-type: none"><li>• Impermeable to blood and body fluids.</li><li>• Single use.</li><li>• Avoid culturally unacceptable colors e.g. black</li><li>• Hot Seam Seal</li><li>• Light colors are preferable to better detect possible contamination</li><li>• Thumb/finger loops to anchors sleeves in place</li><li>• Quality compliant with following standard</li></ul> <p>a) Meets or exceeds ISO 16603 class 3 exposure pressure, or equivalent. b) Material used for manufacturing coverall shall not cause irritation to the user. c) Size universal fit.</p>	<ul style="list-style-type: none"><li>• Impermeable to blood and body fluids.</li><li>• Single use.</li><li>• Avoid culturally unacceptable colors e.g. black</li><li>• Hot Seam Seal</li><li>• Light colors are preferable to better detect possible contamination</li><li>• Thumb/finger loops to anchors sleeves in place</li><li>• Mass of the fabric (Overall Including lamination, Gram per square meter)- 70 to 75 GSM.</li><li>• Quality compliant with following standard</li></ul> <p>a) Meets or exceeds ISO 16603 class 3 exposure pressure, or equivalent. b) Material used for manufacturing coverall shall not cause irritation to the user. c) Size universal fit.</p>

II. Quantity of PPE KIT Amended as (6,00,000+6,00,000) Total**12,00,000.**

4. Quotation Notification No. HFW/KSMSCL/COVID/T-Layer/08/2021-22,  
Date: 27/04/2021.


i. Quantity of Triple Layer Mask Amended as (20,00,000+50,00,000)  
Total **70,00,000.**


5. Quotation Notification No. HFW/KSMSCL/GH-IFQ/06/2021-22,

Dated: 26.04.2021.

- i. Quantity of Hand Sanitizer 1x500ml Amended as (31,500+5,00,000)  
Total **5,31,500**. And per 1x5000ml amended as (3150+1,00,000)  
Total **1,03,150**.

  
Chief Pharmacist  
KSMSCL,  
Bangalore

  
General Manager(Drugs)  
KSMSCL, Bangalore

  
Director(Drugs)  
KSMSCL, Bangalore

  
Managing Director,  
KSMSCL, Bangalore.