



GOVERNMENT OF KARNATAKA

KARNATAKA STATE DRUG LOGISTICS & WAREHOUSING SOCIETY®

No.1, Dr. Siddaiah Puranik Road, KHB Colony, Magadi Road, Bangalore – 560079

Phone: +91-80-23486221 Fax: +91-80-23281477

Email: helpdesk.kdlws@gmail.com, ad.kdlws@gmail.com

No. KDL/EQPT/Covid19/130/2019-20

22-05-2020

CORRIGENDUM-2

INVITATION FOR QUOTATIONS (IFQ) FOR SUPPLY OF FINGER TIP PULSE OXIMETER.

Sl. No	Name of the Items	Schedule of Events	Present Schedule	Postponed / Revised Schedule
1	Finger Tip Pulse Oximeter	Last date and time for Submission of Quotation	22/05/2020 up to 5.00 PM	26/05/2020 up to 05:00 PM

CHANGES IN TECHNICAL SPECIFICATION:

Sl. No	Present Technical Specification	Revised Technical Specifications
1	<ul style="list-style-type: none">➤ Application Range<ul style="list-style-type: none">a. Adult and Paediatrics➤ Measuring Range: 30bpm ~ 240bpm➤ Measuring Accuracy: ± 2bpm or: $\pm 2\%$ whichever is greater Default Alarm Limit SpO2 Low Limit: 90% Pulse Rate: High Limit: 120bpm Low Limit: 50bpm	<ol style="list-style-type: none">1. Display: OLED displays - Oxygen Saturation (%SpO₂), Pulse rate (beats per minute), Respiratory rate (breaths per minute), Plethysmograph, Battery indication. Oxygen saturation: 35% to 100%; Pulse rate display range : 25 bpm to 250 bpm2. Environment condition: Working Temperature: 5°C to 40°C Relative Humidity : $\leq 80\%$, Non-condensing3. Resolution : Oxygen saturation(SPO₂)% - 1% Pulse rate : 1bpm4. Measurement accuracy : Oxygen saturation(SPO₂): ± 2 (70 % to 100%) ; $\leq 70\%$ Unspecified Pulse rate : ± 2 bpm5. Alarm default values: Oxygen saturation: Upper limit: 100 ; Lower limit: 94

Sl. No	Present Technical Specification	Revised Technical Specifications
		Pulse rate: Upper limit: 130 ; Lower limit: 50 6. Physical Characteristics: Dimensions: 74mm x 37mm x 38mm (±3mm) Weight: 55g (±2g) 7. Battery: 2 AAA batteries

Manjunath 22/5/20
Commissioner
Food Safety and Standard Authority
and Additional Director,
KSDLWS Bengaluru